



Product Service

Attestation of Conformity

No. M8A 117198 0001 Rev. 00

Holder of Certificate: Evacle Pte Ltd

22 Sin Ming Lane #06-76 Midview City
Singapore 573969
SINGAPORE

Product:

**Electrical Machinery
(Medical Waste Disinfection System)**

Model(s):

EVA 80 FC

Parameters:

Rated Input Voltage : 380-400VAC, 3P + N + PE
Rated Frequency : 50/60 Hz
Rated Input Current : 32 A
Protection Class : I

**Tested
according to:**

EN ISO 12100:2010
EN 60204-1:2018

This Attestation of Conformity is issued on a voluntary basis according to Council Directive 2006/42/EC relating to machinery. It confirms that the listed equipment (not Annex IV equipment) complies with the principal protection requirements of the directive. It refers only to the sample submitted to TÜV SÜD Product Service GmbH for testing and certification. For details see: www.tuvsud.com/ps-cert

Test report no.:

7191281985-EEC22-TZ

Date, 2022-06-22

(QI THOMAS LUO)